



PFA-860 Footwear Auditor

The Prostat PFA-860 Footwear auditor is designed for resistance measurements of footwear while being worn by personnel. It is designed for use in conjunction with the Prostat PRS-801 and PRS-812 Resistance Meters. The PFA-860 can be used to: (1) measure individual body resistance through the body to the test plate; (2) measure resistance through the body and shoes (one or both shoes) to the test plate; and, (3) measure resistance through a shoe to the test plate.

Specifications

Test Plate:	Two stainless steel plates 1/16 inch thick, welded to ¼ inch stainless steel piano hinge. Equipped with 2 snap connections at opposite corners. Bottom surface laminated with polypropylene insulating material.
Size:	17" x 11" (43 cm x 28 cm). Folds to 8" x 11" x 0.25" (20.3 cm x 27.9 cm x 0.6 cm)
Test Wand:	Stainless steel tube assembly equipped with an insulated sealing cap on end and a banana receptacle for meter lead connection at the opposite end.
Wand Size:	5 ½ inches long with ⅞ inch outer diameter
Banana Snap:	Equipped with an insulated snap cover and lug type banana receptacle.

©2007 Prostat Corporation. All rights reserved.
Prostat is the registered trademarks of Prostat Corporation in the United States and other countries.

1072 Tower Lane, Bensenville, IL 60106 USA
Phone +1 (630) 238-8883 Fax +1 (630) 238-9717
Web access: <http://www.prostatcorp.com>

Specifications subject to change without notice.
Printed in U.S.A. REV2: 6/20/07

1 Prostat Corporation PFA-860 Footwear Auditor



Distributed by

Pro-Pack Materials Pte Ltd

The ESD & Cleanroom Specialist

📍 No. 2, Bukit Batok Street 23, #08-05 Bukit Batok Connection, Singapore 659554. ✉ sales@pro-pack.com.sg ☎ (65) 6354 1511